GM Titanium Base

Single-unit screw-retained prosthesis

Single-unit cement-retained prosthesis

Ø 3.5/4.5/ 5.5/6.5 mm

Check it out on the eShop, go to: neodent.com/shopnow



Installation Sequence

Intraoral



GM Implant Intraoral Scanbody (2) 108.207

GM Implant Analog

Ø 3.5/3.75	Ø 4.0/4.3	Ø 5.0/6.0
101.103	101.089	101.090

Hybrid Repositionable (conventional/digital)

Model Scanning



GM Implant Exact Impression Coping Closed and Open Tray (2)

Regular	108.160	108.162
Long	108.161	108.163



GM Implant Analog

Ø 3.5/3.75	Ø 4.0/4.3	Ø 5.0/6.0
101.103	101.089	101.090

Hybrid Repositionable (conventional/digital)



GM Implant Intraoral Scanbody (2) 108.207

Conventional



GM Implant Exact Impression Coping Closed and Open Tray (2)

Regular	108.160	108.162
Long	108.161	108.163



GM Implant Analog

Ø 3.5/3.75	Ø 4.0/4.3	Ø 5.0/6.0
101.103	101.089	101.090

Hybrid Repositionable (conventional/digital)

GM Exact Titanium Base with Removable Screw 4mm

	0.8 mm	1.5 mm	2.5 mm	3.5 mm	4.5 mm
Ø 3.5	135.355	135.356	135.357	135.358	135.359
Ø 4.5	135.367	135.368	135.369	135.370	135.371
Ø 5.5	135.379	135.380	135.381	135.382	135.383
Ø 6.5		135.391	135.392	135.393	135.394

20 N.cm



GM Exact Titanium Base with Removable Screw 6mm

	0.8 mm	1.5 mm	2.5 mm	3.5 mm	4.5 mm
Ø 3.5	135.361	135.362	135.363	135.364	135.365
Ø 4.5	135.373	135.374	135.375	135.376	135.377
Ø 5.5	135.385	135.386	135.387	135.388	135.389
Ø 6.5		135.395	135.396	135.397	135.398

20 N.cm



GM Titanium Base Burn-out Coping



Ø 3.5	Ø 4.5	Ø 5.5	
118.322	118.325	118.329	4.0 mm
118.323	118.327	118.342	6.0 mm

Drivers

1. Neo Screwdriver Torque Connection + Torque Wrench

2. Neo Screwdriver Torque Connection + Manual Screwdriver Torque

Accessories



Replacement Abutment Screw

116.292 Neo GM Screw (Long)